

Table 2. Resistance to Blood Penetration (ASTM F 1862)

Face Mask  
 Brilliance Air,  
 FM3P-10Pk  
 Brilliance Air, MFR:  
 FEB2021, UVC  
 Sanitized, Made in  
 USA (JN 23834)



AT-1659

In Scope of ISO 17025 Accreditation (Y/N) Y  
 Resistance to Blood Penetration (Pass/Fail)

1	Pass
2	Pass
3	Pass
4	Pass
5	Pass
6	Pass
7	Pass
8	Pass
9	Pass
10	Pass
11	Pass
12	Pass
13	Pass
14	Pass
15	Pass
16	Pass
17	Pass
18	Pass
19	Pass
20	Pass
21	Pass
22	Pass
23	Pass
24	Pass
25	Pass
26	Pass
27	Pass
28	Pass
29	Pass
30	Pass
31	Pass
32	Pass

Resistance to Blood Penetration Summary 32 Pass/0 Fail

Table 2. Resistance to Blood Penetration (ASTM F 1862) (contd.)

---

	Face Mask Brilliance Air, FM3P-10Pk Brilliance Air, MFR: FEB2021, UVC Sanitized, Made in USA (JN 23834)
<hr/>	
Test Parameters	
Synthetic Blood Lot #	109201
Test Pressure (mmHg)	160

---

Method(s) and Notes:

All valid results are included in the statistical analyses.  
Revisions of SGS-IPS methods when used are current at the time of testing.  
Samples tested and conditioned in TAPPI standard conditions unless requested otherwise by customer or otherwise specified.  
Samples were not preconditioned.  
16 CFR Ch. II Part 1610 (1-1-18 Edition) - Standard for the Flammability of Clothing Textiles  
Testing conducted as a Plain Surface Textile Fabric.  
Samples were not refurbished.  
Preliminary tests to determine most rapidly burning direction were not conducted.  
Samples were cooled in a desiccator after oven drying for a minimum of 15 minutes.  
If samples could not be dried in an oven, samples were placed in a desiccator for a minimum of 1 hour.  
Specimens were cut to 2.5"x6" to assist in securing specimens in the specimen holder.  
ASTM F 1862/F 1862M-17 Standard Test Method for Resistance of Medical Face Masks to Penetration by Synthetic Blood (Horizontal Projection of Fixed Volume at a Known Velocity)  
A 2 mL check was performed at the beginning, middle and end of each sample.  
Synthetic blood is purchased from Johnson, Moen & Co. Surface Tension is not independently verified after receipt and unused synthetic blood is stored in original plastic bottles.  
Samples were conditioned prior to testing. Chamber conditions can be found as reported.

Analyzed by: LAH, SW  
Quality review by: MAG  
Date(s) of testing: February 24-25, 2021



**Room Conditions**

	Relative Humidity (%)	Temperature (°F)
Conditioning Environment	49.6	73.6
Maximum during testing	49.9	73.6
Minimum during testing	49.6	73.2

**Chamber Conditions ASTM F 1862 160mmHg**

	Relative Humidity (%)	Temperature (°F)
Chamber Maximum	83.6	72.3
Chamber Minimum	80.1	71.7

Note: See the method(s) cited above for available estimates of measurement uncertainty. Unless otherwise noted, sampling was performed by customer.

This document is issued by the Company with the expectation that the document will be used internally only and not be used for any marketing or any other public use and is subject to its General Conditions of Service printed overleaf, available on request or accessible at [www.sgs.com/terms\\_and\\_conditions.htm](http://www.sgs.com/terms_and_conditions.htm) and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at [www.sgs.com/en/Terms-and-Conditions/terms-e-document.aspx](http://www.sgs.com/en/Terms-and-Conditions/terms-e-document.aspx). Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.